

**United States Department of Commerce
National Institute of Standards and Technology**

**~ Certificate of Measurement Traceability ~
for**

Nebraska

**a participant in the NIST Weights and Measures
Measurement Assurance Program for the Year(s)**

2006

**The State standards are traceable to NIST and the laboratory
has *conditionally* demonstrated* measurement proficiency
through training and interlaboratory comparisons.**



*Georgia L. Harris,
for the National Institute of Standards and Technology*

Laboratory Metrology, Group Leader, Weights and Measures Division
Title

January 13, 2006

Effective Date

***Demonstrated
Areas of Proficiency**

Mass

Echelon II
10 kg - 1 mg
Echelon III
2500 lb - 0.001 lb

Volume

Gravimetric
1 gal - 2 fl oz
Volume transfer
100 gal - 5 gal

*Conditional Status
Improvements to the quality sys-
tem documentation are needed
and an on-site assessment is
needed.*

